

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 02 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130114</u> ✓
Name <u>Kevin Stevenson</u>		ID# <u>2414</u> Date <u>05-29-09</u> ✓
A	Agency <u>UAA Police Dept</u>	Phone # <u>907-786-1120</u>
Instrument Location <u>UAA Police Dept</u> ✓		
Alco S/N <u>75470</u> ✓		Target Value <u>.080</u> ✓ High Pressure <u>1100</u>
B	Alco Test Value Average <u>.081</u> ✓ 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco <u>.081</u> ✓	
Signature <u>[Signature]</u> ✓		
(OVER)		

BMB  
6/3/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 4<sup>th</sup> day of June, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114 ✓  
MAY 29, 2009 ✓ *BMG 6/3/09*  
OPERATOR'S NAME: STEVENSON KEVIN L.  
OPERATOR'S NUMBER: 2414  
SUBJECT'S LAST NAME: VERIFICATION  
SUBJECT'S FIRST NAME/NI: VERIFICATION  
O.L. #: 09999  
DEPT/AGENCY: RNCZ  
CASE/REPORT: 05292009  
TEST TYPE: U  
ALCO TARGET VALUE: .000 ✓  
ALCO S/N: 75470 ✓  
— BREATH ANALYSIS —  
.000 ADJUSTED FOR 29.86 in  
ALCO TARGET .079 18:48  
BLANK TEST .000 18:48  
INTERNAL STANDARD VERIFIED 18:49  
ALCO TV 29.86 in .081 ✓ 18:49  
BLANK TEST .000 18:50  
SUBJECT SAMPLE *BMG 6/3/09* .000 18:50  
BLANK TEST .000 18:51  
ALCO TV 29.86 in .082 18:51  
BLANK TEST .000 18:53

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130114  
MAY 29, 2009 ✓  
TIME 18:44  
— DIAGNOSTIC CHECK —  
COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.86 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY  
PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOPQRSTUVWXYZ\]^\_`ab cdefghijklmno  
pqrstuvwxyz{|}~